

Serveur Académique Lausannois SERVAL serval.unil.ch

Author Manuscript

Faculty of Biology and Medicine Publication

This paper has been peer-reviewed but does not include the final publisher proof-corrections or journal pagination.

Published in final edited form as:

Title: Performance of an automated multiplex immunofluorescence assay for detection of *Chlamydia trachomatis* immunoglobulin G.

Authors: Baud D, Zufferey J, Hohlfeld P, Greub G

Journal: Diagnostic microbiology and infectious disease

Year: 2014 Mar

Volume: 78

Issue: 3

Pages: 217-9

DOI: [10.1016/j.diagmicrobio.2013.11.022](https://doi.org/10.1016/j.diagmicrobio.2013.11.022)

In the absence of a copyright statement, users should assume that standard copyright protection applies, unless the article contains an explicit statement to the contrary. In case of doubt, contact the journal publisher to verify the copyright status of an article.

**PERFORMANCE OF AN AUTOMATED MULTIPLEX IMMUNOFLUORESCENCE
ASSAY FOR DETECTION OF *CHLAMYDIA TRACHOMATIS* IMMUNOGLOBULIN G**

David Baud^{1,2}, Jade Zufferey², Patrick Hohlfeld², Gilbert Greub^{1*}

¹ Center for Research on Intracellular Bacteria, Institute of Microbiology, Faculty of Biology and Medicine, University of Lausanne and University Hospital Center (CHUV), Lausanne, Switzerland

² Materno-fetal and Obstetrics Research Unit, Departments of Obstetrics and Gynecology, University Hospital Center (CHUV), Lausanne, Switzerland

***CORRESPONDING AUTHOR:** Gilbert Greub

Center for Research on Intracellular Bacteria (CRIB)

Institute of Microbiology - University of Lausanne

Bugnon 48

1011 Lausanne - SWITZERLAND

Phone: (00) 41 21 314 49 79

Fax: (00) 41 21 314 40 60

Email: gilbert.greub@chuv.ch

RUNNING TITLE:

Chlamydia trachomatis & serological tests

ABSTRACT

Chlamydia serology is indicated to investigate etiology of miscarriage, infertility, PID and ectopic pregnancy. Here, we assessed the reliability of a new automated-multiplex immunofluorescence assay (InoDiag test) to detect specific anti-*C. trachomatis* IgG. Considering IF as gold standard, InoDiag tests exhibited similar sensitivities (65.5%), but better specificities (95.1%-98%) than ELISAs. InoDiag tests demonstrated similar or lower cross-reactivity rates when compared to ELISA or IF.

KEY WORDS:

Chlamydia trachomatis, intracellular bacteria, serology, diagnostic test

INTRODUCTION

Prevalence of *Chlamydia trachomatis* genital infections is steadily increasing over the last decade (Rekart et al., 2013). The majority of chlamydial infections remain asymptomatic and therefore undetected causing a worldwide silent epidemic (Paavonen, 2012). Left untreated, *Chlamydia trachomatis* may lead to ascending infection and tubal damage with chronic sequelae such as tubal infertility and sterility (Baud and Greub, 2011; Paavonen, 2012). Moreover, *Chlamydia trachomatis* is recognised as an agent of miscarriage (Baud et al., 2011). Although PCR represents the ideal diagnostic approach for urethritis and cervicitis, serology remains useful to detect anti-chlamydial antibodies among women with tubal infertility, ectopic pregnancy, miscarriage, pelvic inflammatory disease and chronic pelvic pain (Baud et al., 2008; Baud et al., 2011; Baud and Greub, 2011; Haggerty et al., 2010), since in such chronic conditions, the negative predictive value is very high. Microimmunofluorescence assay (IF), still considered as the “gold standard” for the serology of this pathogen, has the disadvantage to be time-consuming, subjective and reader-dependant (Gaydos C and Essig A, 2011). However, enzyme-linked immunosorbent assays (ELISA) using the major outer membrane protein (MOMP) as antigen appears to exhibit a higher sensitivity for diagnosing chlamydial infections (Baud et al., 2010).

The present study aimed to investigate the performance of a new fully automatized “multiplex fluorescence immuno-assay” (Inodiag[®], Signes, France) to detect specific anti-*C. trachomatis* IgG. Two different assay formats were investigated: Inodiag-EBs based on antibody reactivity to the elementary body (EB) and Inodiag-MOMP based on the MOMP of chlamydiae. Both Inodiag tests were compared with two other commercialized MOMP ELISA tests (MOMP-Medac[®] and MOMP-RBiopharm[®]) and to IF, which was considered to be the the gold standard.

METHOD

Studied sera were obtained from women with and without miscarriages as described previously by Baud *et al.* (Baud *et al.*, 2007; Baud *et al.*, 2009; Baud *et al.*, 2010). A total of 265 sera were tested for the presence of *Chlamydia trachomatis* IgG antibodies with the new automated “multiplexed serology test” (Mu.S.T) as previously described (Baud *et al.*, 2010; Gouriet *et al.*, 2008a; Gouriet *et al.*, 2008b), but using a new InoDiag platform allowing simultaneous testing of up to 12 patients in a single run. Practically, 2 different *Chlamydia trachomatis* antigens are present in the test: elementary bodies (InoDiag-EB) and MOMP (InoDiag-MOMP). Results of EBs and MOMP antigens were analysed separately and combined (hereafter called InoDiag-Combi, see details in Table 1).

Inodiag tests were compared with 3 other serological tests: immunofluorescence (Micro IF, ANILabsystems, Vantaa, Finland) and two *C. trachomatis* IgG ELISA both using the MOMP as antigen: MOMP-Medac, CT-IgG-pELISA, Medac, Webel, Germany and MOMP-R, CT pELISA, R-Biopharm, Darmstadt, Germany. All these 3 commercial tests were performed according to the manufacturers’ instructions.

To evaluate cross-reactivity, serum samples collected from patients with positive serologies against *Toxoplasma gondii* (n=81 cases), *Waddlia chondrophila* (n=83), *Chlamydia psittaci* (n=24), *Chlamydia pneumoniae* (n=96) and *Brucella* sp (n=11) were also analysed and the risk of cross-reactivity with *Chlamydia trachomatis* was statistically determined. All these serological tests were performed using commercial kits previously reported (Baud *et al.*, 2007, Baud *et al.*, 2010) and for *Waddlia chondrophila* using an home made immunofluorescence as earlier described (Baud *et al.*, 2007).

To compare all five serological tests, the sensitivities, specificities, positive predictive values (PPVs) and negative predictive values (NPVs) were calculated using Stata (StataCorp, College Station, TX, USA). Values in the grey zone (i.e. with optical densities between the values

threshold for negativity and positivity, as defined by the manufacturer of MOMP-Medac and MOMP-RB) were excluded from these analyses. To assess cross-reactivity, we compared the correlation between serology directed against two different pathogens using the Chi² test.

RESULTS

Of the 265 sera, *Chlamydia trachomatis* IgG seropositivity rate was 14%, 19%, 16%, 11%, 13% and 9% with IF, MOMP-Medac, MOMP-RB, InoDiag-EB, InoDiag-MOMP and InoDiag-Combi, respectively. Inconclusive results were respectively observed in 0%, 3%, 7%, 2%, 3% and 10% of these serological tests. The level of agreement between the assays was similar. The percentage of concordance with results of IF was 63.9% for MOMP-Medac, 61.1% for MOMP-RBiopharm, 63.9% for Inodiag-EBs, 63.9% for MOMP and 55.6 % for Inodiag-Combi.

Table 1 shows performance of the different assay using IF as gold standard. A total of 33 inconclusive samples (12%) were excluded from the inter-assays comparison as they exhibited a doubtful result by at least one assay. MOMP-RB and InoDiag-Combi exhibited the best sensitivities (72.4%), whereas InoDiag-EB exhibited the best specificity (98%).

A second analysis was conducted using a modified gold standard based on the results from IF, MOMP-Medac and MOMP-RB (see footnotes of Table 1 for details). This analysis involved 197 samples including 17 *Chlamydia trachomatis*-positive and 177 *Chlamydia trachomatis*-negative samples. InoDiag-MOMP and InoDiag-Combi exhibited the best sensitivities (100%), whereas InoDiag-EB exhibited the best specificity (99.4%).

All 6 tests cross-reacted with *C. psittaci* ($p < 0.001$) and *C. pneumoniae* ($p < 0.05 - 0.001$) as shown in Table 2. Moreover, IF also cross-reacted significantly with *T. gondii* and *W. chondrophila* (Table 1). InoDiag-MOMP showed the lowest cross-reactivity rate, with the different investigated species.

DISCUSSION

In this study, we assessed the performance of an automated multiplex antigen microarray (InoDiag) for the detection of *Chlamydia trachomatis* IgG antibodies.

The specificity of the Inodiag assays (94.6%-98%) was higher than the two commercialized ELISA used for comparison, whereas sensitivities were similar (65.5%-72.4%). In the present study, we showed that sensitivity was not improved by MOMP alone compared to EBs. However, the combination of MOMP and EBs by Inodiag increased sensitivity without altering specificity. All the tests studied here show low PPV, indicating that confirmation is required when positive as suggested by others (Mylonas, 2012). Moreover, we would recommend to use a second serological assay in case of doubtful result (grey zone).

The “gold standard” IF exhibits significant limitations: (i) high levels of cross-reactions with other members of the *Chlamydiaceae* family, (ii) labour-intensive and (iii) subjective and operator-dependant reading. In contrast, ELISAs (Medac or R-Biopharm) represent a test characterized by the following advantages: (i) objective reading of the results, (ii) less expensive than IF (especially when considering technician time) and (iii) high throughput, useful for epidemiological studies.

In contrast to our study, Muvunyi et al. showed that peptide-based ELISA performed as well as two different IF assays (Muvunyi et al, 2012). However, the performance of the InoDiag tests presented here are consistent with previous studies that revealed good sensitivities and specificities of ELISA assays based on peptides from the MOMP (Muvunyi et al, 2012 ; Land JA et al, 2003). The new automated microarray from Inodiag® show similar performance than ELISA. It is well designed for small laboratories, since batches of only 2 to 4 samples can be tested with InoDiag. Several pathogens can be tested simultaneously, allowing pathology-driven testing instead of the common pathogen-driven testing (Gouriet et al., 2008a; Gouriet et al., 2008b; Raoult et al., 2004). Moreover, thank to the multiplex format of the Inodiag technology,

new markers of chronic *C. trachomatis* infection might be added to improve serodiagnosis (Cappello et al., 2009).

ACKNOWLEDGEMENTS

We thank Sébastien Aeby and Joel Gyger for the technical help, Karine Lepigeon and Françoise Damnon for computer assistance. David Baud is supported by the “Fondation Leenaards” through the “Bourse pour la relève académique”, the “Société Académique Vaudoise” through the “Paul Blanc” grant, the SICPA Foundation and an Air Canada Travel Grant.

Reference List

- Baud,D., Goy,G., Jaton,K., Osterheld,M.C., Blumer,S., Borel,N. et al. (2011)** Role of *Chlamydia trachomatis* in miscarriage. *Emerg Infect Dis* **17**: 1630-1635.
- Baud,D. and Greub,G. (2011)** Intracellular bacteria and adverse pregnancy outcomes. *Clin Microbiol Infect* **17**: 1312-1322.
- Baud,D., Kebbi,C., Kulling,J.P., and Greub,G. (2009)** Seroprevalence of different *Chlamydia*-like organisms in an asymptomatic population. *Clin Microbiol Infect*.
- Baud,D., Regan,L., and Greub,G. (2008)** Emerging role of *Chlamydia* and *Chlamydia*-like organisms in adverse pregnancy outcomes. *Curr Opin Infect Dis* **21**: 70-76.
- Baud,D., Regan,L., and Greub,G. (2010)** Comparison of five commercial serological tests for the detection of anti-*Chlamydia trachomatis* antibodies. *Eur J Clin Microbiol Infect Dis* **29**: 669-675.
- Baud,D., Thomas,V., Arafa,A., Regan,L., and Greub,G. (2007)** *Waddlia chondrophila*, a potential agent of human fetal death. *Emerg Infect Dis* **13**: 1239-1243.
- Cappello,F., Conway de,M.E., Di,F., V, Zummo,G., and Macario,A.J. (2009)** *Chlamydia trachomatis* infection and anti-Hsp60 immunity: the two sides of the coin. *PLoS Pathog* **5**: e1000552.
- Gaydos C and Essig A 2011.** Chlamydiaceae, Manual of Clinical Microbiology. AMS Press, Washington DC. [10th Edition], p986-1000.

Geisler,W.M., Morrison,S.G., Doemland,M.L., Iqbal,S.M., Su,J., Mancevski,A. et al. (2012) Immunoglobulin-specific responses to *Chlamydia* elementary bodies in individuals with and at risk for genital chlamydial infection. *J Infect Dis* **206**: 1836-1843.

Gouriet,F., Levy,P.Y., Samson,L., Drancourt,M., and Raoult,D. (2008a) Comparison of the new InoDiag automated fluorescence multiplexed antigen microarray to the reference technique in the serodiagnosis of atypical bacterial pneumonia. *Clin Microbiol Infect* **14**: 1119-1127.

Gouriet,F., Samson,L., Delaage,M., Mainardi,J.L., Meconi,S., Drancourt,M., and Raoult,D. (2008b) Multiplexed whole bacterial antigen microarray, a new format for the automation of serodiagnosis: the culture-negative endocarditis paradigm. *Clin Microbiol Infect* **14**: 1112-1118.

Haggerty,C.L., Gottlieb,S.L., Taylor,B.D., Low,N., Xu,F., and Ness,R.B. (2010) Risk of sequelae after *Chlamydia trachomatis* genital infection in women. *J Infect Dis* **201 Suppl 2**: S134-S155.

Land JA, Gijsen AP, Kessels AG, Slobbe ME, Bruggeman CA. (2003) Performance of five serological *chlamydia* antibody tests in subfertile women. *Hum Reprod.* 18(12):2621-7.

Molano,M., Meijer,C.J., Weiderpass,E., Arslan,A., Posso,H., Franceschi,S. et al. (2005) The natural course of *Chlamydia trachomatis* infection in asymptomatic Colombian women: a 5-year follow-up study. *J Infect Dis* **191**: 907-916.

Muvunyi CM, Claeys L, De Sutter T, De Sutter P, Temmerman M, Van Renterghem L, Claeys G, Padalko E. (2012) Comparison of four serological assays for the diagnosis of *Chlamydia trachomatis* in subfertile women. *J Infect Dev Ctries.* 14;6(5):396-402.

Mylonas,I. (2012) Female genital *Chlamydia trachomatis* infection: where are we heading? *Arch Gynecol Obstet* **285**: 1271-1285.

Paavonen,J. (2012) *Chlamydia trachomatis* infections of the female genital tract: state of the art. *Ann Med* **44**: 18-28.

Raoult,D., Fournier,P.E., and Drancourt,M. (2004) What does the future hold for clinical microbiology? *Nat Rev Microbiol* **2**: 151-159.

Rekart,M.L., Gilbert,M., Meza,R., Kim,P.H., Chang,M., Money,D.M., and Brunham,R.C. (2013) *Chlamydia* public health programs and the epidemiology of pelvic inflammatory disease and ectopic pregnancy. *J Infect Dis* **207**: 30-38.

1 **Table 1-bis:** Performance of the 3 INODIAG serological tests using a new gold standard⁽¹⁾.

2 The 95% confidence intervals are shown in brackets.

3

Assay	Sensitivity	Specificity	PPV	NPV
INODIAG - EBs	89.5% [66.9-98.7]	99.4% [96.9-100]	94.4%[72.7-99.9]	98.9% [96-99.9]
INODIAG-MOMP	100% [82.4-100]	97.2% [93.5-99.1]	79.2% [57.9-92.9]	100% [97.9-100]
INODIAG-Combi*	100% [82.4-100]	96.6% [92.8-98.8]	76% [54.9-90.6]	100% [97.9-100]

4

5

6 *InoDiag-Combi was designed according to the results of InoDiag-EB and InoDiag-MOMP.

7 True positives were defined as both tests positive and true negatives as both tests negative.

8 Discordant results or values in the grey zone were excluded from the analysis.

9

10 ⁽¹⁾ Performance of the tests using a new gold standard based on the results of IF, MOMP-Medac
11 and MOMP-RB. True positives were defined as all 3 tests positive and true negatives as all 3
12 tests negative. All discordant results or results in the grey zone were excluded from the analysis.

1 **Table 2:** Rate of *Chlamydia trachomatis* positivity of each serological assay for sera
 2 identified IgG-positive for another pathogenic agent.

3

Assay	<i>Toxoplasma gondii</i> n=81	<i>Waddlia chondrophila</i> n=83	<i>Chlamydia psittaci</i> n=24	<i>Chlamydia pneumoniae</i> n=96	<i>Brucella abortus</i> n=11
IF	21% [12-30] (+)	7% [2-13] (+)	79% [62-97] (+++)	27% [18-36] (+++)	9% [0-29]
MOMP-Medac	15% [7-23]	16% [8-24]	67% [46-87] (+++)	29% [20-38] (++)	9% [0-29]
MOMP-RBiopharm	14% [6-21]	13% [6-21]	71% [51-90] (+++)	30% [21-40] (+++)	9% [0-29]
INODIAG - EBs	12% [5-20]	6% [1-11]	79% [62-97] (+++)	20% [12-28] (++)	0% [0-0]
INODIAG-MOMP	11% [4-18]	11% [4-18]	67% [46-87] (+++)	20% [12-28] (+)	0% [0-0]
INODIAG-Combi	9% [2-15]	6% [1-12]	67% [46-87] (+++)	16% [8-23] (++)	0% [0-0]

4

5 p-value < 0.05 = (+)

6 p-value < 0.01 = (++)

7 p-value < 0.001 = (+++)